



IMMUNOWELL™

CARDIOLIPIN ANTIBODY (IGG) TEST

Product No. 3090



For In Vitro Diagnostic Use

SEE CALIBRATION VALUES, TABLE 2, PAGE 3

INTENDED USE

The ImmunoWELL Cardiolipin Antibody (IgG) Test is an enzyme immunoassay (EIA) for the quantitative detection of antibodies to cardiolipin antigen in serum and is used as an aid for assessing the risk of thrombosis in individuals with systemic lupus erythematosus (SLE) and lupus-like disorders.

SUMMARY AND EXPLANATION

Anticardiolipin (aCL) autoimmune antibodies are a subset of antibodies which react with negatively charged phospholipids. One type of anti-phospholipid, lupus anticoagulant (LA), is frequently found in systemic lupus erythematosus (SLE) patients. LA is defined as an immunoglobulin that prolongs in vitro coagulation tests (e.g., APTT or PT). Anticardiolipin antibodies are also frequently observed in patients with SLE. One autoantibody, LA, is measured with a bioassay and the other autoantibody, aCL, is measured based on affinity for cardiolipin. In many cases both autoantibodies are present in a single patient and may be the same antibody in some cases. Thrombosis and spontaneous abortion are among the symptoms common to SLE patients; in fact, elevated levels of anticardiolipin antibodies have been strongly associated with the presence of both venous and arterial thrombosis, thrombocytopenia, and recurrent fetal loss. (1) (2) Patients who present with these manifestations are known to have an "anti-phospholipid syndrome". (2) (3) (4) (5) (6)

Anticardiolipin (aCL) and lupus anticoagulant (LA) are not limited to SLE patients. In fact, the majority of people who have these autoantibodies do not have SLE. These antibodies occur in patients with other autoimmune diseases and in individuals with no apparent autoimmune disease (3) (7) (8); yet, many subjects manifest one or more of the clinical complications associated with the presence of anti-phospholipid antibodies in their circulation.

The ImmunoWELL Cardiolipin Antibody (IgG) Test is an EIA which measures IgG aCL in human serum. The ImmunoWELL test provides highly reproducible results expressed in units that are standardized against an internationally recognized reference preparation. Levels of IgG aCL antibodies are reported as GPL units. One GPL unit is defined as the cardiolipin binding activity of 1 µg/ml of an affinity purified IgG aCL preparation from a standard serum (REY) (9).

ASSAY PRINCIPLE

The ImmunoWELL Test utilizes an EIA microtiter plate technique for the detection of antibodies. Serum is added to antigen coated microtiter wells and allowed to react. After removal of unbound antibodies, horseradish peroxidase-conjugated antihuman antibodies are allowed to react with bound antibodies. The bound peroxidase reacts with 2,2'-azino-di-[3-ethylbenzthiazoline sulfonate] (ABTS®), the chromogenic substrate, developing a color. Finally, the substrate reaction is stopped and the optical density is read with a spectrophotometric microwell reader.

REAGENTS

Reaction Wells coated with diphosphatidyl glycerol (cardiolipin from bovine heart).

Specimen Diluent: Consisting of 0.01M phosphate buffered saline (PBS, pH 6.2-7.6) and 10% newborn calf serum containing <0.1% NaN₃.

Positive Control: Consisting of human anti-cardiolipin serum containing <0.1% NaN₃.

Negative Control: Consisting of nonreactive human serum containing <0.1% NaN₃.

Calibrators (3): Consisting of human anti-cardiolipin sera at specified concentrations containing <0.1% NaN₃.

Wash Buffer Concentrate: Consisting of a 20X concentrate of 0.01M PBS (pH 6.2-7.6)

Conjugate: Consisting of peroxidase-conjugated goat antihuman IgG (H+L) in PBS (pH 6.2-7.6) and carrier protein containing preservative

Substrate Buffer consisting of 0.1 M sodium citrate (pH 4.4-4.6) and 0.01% hydrogen peroxide

Substrate Concentrate 2.19% 2,2'-azino-di-[3-ethylbenzthiazoline sulfonate] (ABTS) in 0.1 M sodium citrate (pH 4.4-4.6)

Stop Solution 0.25 M Oxalic Acid

WARNINGS AND PRECAUTIONS

For In Vitro Diagnostic Use: ImmunoWELL reagents have been optimized for use as a system. Do not substitute other manufacturers' reagents or other ImmunoWELL Test reagents. Dilution or adulteration of these reagents may also affect the

performance of the test. Do not use any kits beyond the stated expiration date. Close adherence to the test procedure will assure optimal performance. Do not shorten or lengthen stated incubation times since this may result in poor assay performance.

Sodium azide may react with lead and copper plumbing to form highly explosive metal azides. It may be harmful if enough is ingested (more than supplied in kit). On disposal of liquids, flush with a large volume of water to prevent azide build-up (10). This dilution is not subject to GHS, US HCS and EU Regulation 2008/1272/EC labeling requirements.

0.25M Oxalic acid [$\sim 3\%$ C₂H₂O₄], CAS# 144-62-7, EC No 205-634-3 may be harmful if swallowed so handle according to GLP. This dilution is not subject to GHS, US HCS and EU Regulation 2008/1272/EC labeling requirements.

The safety data sheet (SDS) is available at support.genbio.com or upon request.



Human source material. Material used in the preparation of this product has been tested and found non-reactive for hepatitis B surface antigen (HBsAg), antibodies to hepatitis C virus (HCV), and antibodies to human immunodeficiency virus (HIV-1 and HIV-2). Because no known test method can offer complete assurance that infectious agents are absent, handle reagents and patient samples as if capable of transmitting infectious disease (11). Follow recommended Universal Precautions for bloodborne pathogens as defined by OSHA (12), Biosafety Level 2 guidelines from the current CDC/NIH Biosafety in Microbiological and Biomedical Laboratories (13), WHO Laboratory Biosafety Manual (14), and/or local, regional and national regulations.

RECONSTITUTION AND STORAGE

Kit is stored at 2-8°C. Assuming good laboratory practices are used, opened reagents remain stable as indicated by the expiration date.

Reaction wells are removed from the foil pouch and unused wells are resealed in the pouch using the integral zip-lock.

Wash Buffer (pH 6.2-7.6) is prepared by adding the contents of the Wash Buffer Concentrate (20X) bottle into 1 liter of distilled/deionized water. After reconstitution, the 1X solution is stored at 2-8°C. Discard when visibly turbid.

Note: In some instances the Wash Buffer Concentrate (20X) may develop crystals upon storage at 2-8°C. It is important that these crystals are completely redissolved before dilution of the Concentrate. This can be accomplished by warming the Concentrate to 37°C in a water bath with occasional mixing.

Color Developer is prepared by adding one (1) drop of Substrate Concentrate to 1mL of Substrate Buffer. One mL of Color Developer is sufficient for one eight-well strip. **Use within one hour.**

SPECIMEN COLLECTION AND HANDLING

ImmunoWELL Test is performed on serum. The test requires 10 µL of serum. Serum is collected according to standard practices and may be stored at 2-8° C for up to five days. Serum may be frozen below -20° C for extended periods.

MATERIALS PROVIDED

Microtiter Wells in carrier	Specimen Diluent
Positive Control	Calibrators
Negative Control	Conjugate
Wash Buffer Concentrate (20X)	Substrate Concentrate
Substrate Buffer	Stop Solution

MATERIALS REQUIRED BUT NOT PROVIDED

Distilled or deionized water	Test tubes
Microwell washer	Pipets
Microwell spectrophotometer (405 nm)	

PERFORMANCE CONSIDERATIONS

Reproducibility in the assay is largely dependent upon the consistency with which the microwells are washed. Carefully follow the recommended washing sequence as outlined in the assay procedure.

Positive and Negative Control Sera (Undiluted) are used to assure test performance.

Calibrators (prediluted) are used to construct a standard curve.

Substrate Blank - All reagents, except serum, are added to the substrate blank well. This blank well is intended to baseline (zero) the microwell spectrophotometer.

ASSAY PROCEDURE

1. Allow all components including diluted Wash Buffer to warm to room temperature (22-27°C).
2. Determine the total number of specimens to be tested. Include one blank and duplicates of calibrators and controls in each run. Single test points are adequate for specimen testing, but duplicate tests are recommended for the controls and calibrators.

- For each control and specimen, pipet 10 µL serum into a clean tube containing 1 mL Specimen Diluent and mix (1:100 dilution).

CAUTION: CALIBRATORS ARE PREDILUTED. DO NOT DILUTE FURTHER.

- Determine the total number of wells to be run including blank, calibrators, controls, and specimens. Well strips can be broken to the exact number needed to conserve reagent wells. Strips need to be completed with used wells to facilitate washing procedures.
- Add 100 µL of Specimen Diluent to the first well as a substrate blank. Pipet 100 µL of the prediluted calibrators and the diluted controls and specimens into each assigned well.
- Incubate at room temperature (22-27°C) for 60±2 minutes.
- Aspirate the samples out of the wells.
- Wash the wells three times by completely filling the wells with Wash Buffer (see Reconstitution and Storage) and aspirating the wells completely after washes.
- Pipet 100 µL Conjugate into all wells.
- Incubate the wells at room temperature (22-27°C) for 30±1 minutes.
- Aspirate the conjugate out of the wells.
- Wash the wells three times as described in step 8.
- Prepare fresh Color Developer (see Reconstitution and Storage).
- Pipet 100 µL of Color Developer into each well.
- Incubate at room temperature (22-27°C) for 30±1 minutes.
- Add 100 µL of Stop Solution to each well.
- Inspect the outside bottom surface of the microwells for the presence of condensation, dried buffer salts or wash solution, which might interfere with the spectrophotometric reading. Carefully clean the well bottoms with a soft tissue.
- Using the substrate blank to zero the spectrophotometer, read the optical density of each well at 405 nm within 30 minutes of completion of step 16.

It is recommended that dual wavelength spectrophotometers use only one wavelength, 405 nm.

QUALITY CONTROL

The average of the Positive Control must be within the expected range. Determination of this range is described below in the Interpretation section. Expected Ranges are found in Table 2 of that section. Because test methods may vary between laboratories, it is recommended that the laboratory maintain a record of control values to assure testing consistency and use the expected range as a guideline.

Calibrator values (GPL units) were assigned by simultaneous testing with the standards of Dr. E.N. Harris of the Anti-phospholipid Standardization Laboratory, University of Louisville¹¹.

INTERPRETATION

Construct a point-to-point standard curve using the absorbance values you observe and their corresponding assigned values. Use this curve to calculate antibody concentration of controls and specimens. Levels of aCL representing clinical significance are not well established. There is general agreement that repeated thrombosis or fetal loss occurs in subjects with IgG isotype above 40 GPL units (15). It is recommended that reports be made according to the following:

Table 1: Interpretation

	GPL Units
High Positive	>100
Moderate Positive	>20 – 100
Low Positive	10 – 20
Negative	<10

Table 2: Calibration Values

	Values	Units
High Calibrator Assigned Value	90	Units/mL
Mid Calibrator Assigned Value	51	Units/mL
Low Calibrator Assigned Value	1	Units/mL
Positive Control Expected Range	10-62	Units/mL
Negative Control Expected Range	<10	Units/mL

LIMITATIONS

The anti-cardiolipin activity values obtained from the assay are an aid to diagnosis only. Each physician must interpret these results in light of the patient's history, physical findings, and other diagnostic procedures. If clinical findings suggest the presence of anti-phospholipid antibodies, and the patient is negative for anti-cardiolipin antibodies, some investigators recommend testing for the lupus anticoagulant to confirm the negative result. A patient is considered positive for anti-phospholipid antibodies if one or both of the tests give positive results (16).

Patients with current or prior syphilis infections may have a positive result in the GenBio Cardiolipin Antibody (IgG) Test without risk of thrombosis. If a patient's history is suggestive of syphilis, this possible diagnosis should be ruled out by an assay specific for anti-treponemal antibodies. Anti-cardiolipin antibodies can appear transiently during many infections. If a patient first tests positive while there are clinical signs of infection, the test should be repeated after an interval of 6 months (17).

EXPECTED RESULTS

Although the limit of sensitivity to measure the aCL analyte is 10 GPL units, the limit (cutoff) which measures a level of antibody significantly above a normal, asymptomatic population is higher. This limit was determined by measuring serum in forty-one normal, asymptomatic subjects. The average titer of aCL was 10.6 GPL units with a range between 0 and 52 GPL units and a standard deviation of 9.1 GPL units. These data further support that values above approximately 36 GPL units are significantly different than expected within a normal population.

Thirty-six (36) serum samples from individuals diagnosed as having SLE were tested with the ImmunoWELL Cardiolipin Antibody (IgG) Test. Thirty of the 36 sera were classified as positive based on a 10 GPL cutoff. Eleven patients showed low antibody levels and the remaining 19 had moderate antibody titers. Using an interpretation based on the likelihood of difference from results in normal patients, six of the thirty-six patients had antibody levels above the 36 GPL cutoff (three standard deviations above the mean of normals). Three of these six patients also had significant elevation of the IgM aCL.

PERFORMANCE CHARACTERISTICS

Twenty-five (25) presumptive positive serum samples were tested using the ImmunoWELL Cardiolipin Antibody (IgG) Test and the results were compared to an EIA test performed by a clinical reference laboratory. Both the ImmunoWELL test and the reference lab EIA test normalize data against internationally recognized standards (Antiphospholipid Standardization Laboratory) for anticardiolipin antibodies (aCL). ImmunoWELL performed essentially the same as the reference test method, showing better than 99% correlation. Eight of the 25 samples were above a cutoff of 36 or 40 GPL units in the reference test and were also above 36 GPL in ImmunoWELL (100% sensitivity).

Forty-one (41) normal, asymptomatic sera collected from blood donors were tested with the ImmunoWELL Cardiolipin Antibody (IgG) Test. Using the ASL guidelines described under the Interpretation Section, 19 (46%) of the normal samples were negative, 18 (44%) were identified as low positive, and 4 (10%) were classified as moderate positive. Only one sample had an antibody level (52 GPL) significantly higher than the mean of these normals. The normal serum, which tested positive on the ImmunoWELL test, was run on a second commercial EIA test kit and found to be positive. Based on these results using a cutoff level of 36 GPL units, the overall specificity of the ImmunoWELL Cardiolipin Antibody (IgG) Test is >99% (40/40).

To measure reproducibility, a series of samples were tested in triplicate in five different runs. The results are shown in Table 3.

Table 3: Reproducibility

Mean (GPL)	Within Assay Variation	Between Assay Variation
21	6.0%	14.0%
38	6.2%	13.9%
59	5.9%	8.9%
77	3.7%	12.5%

Variation is presented as Coefficient of Variation (%CV).

BIBLIOGRAPHY

1. *Thromboembolism in patients with the "lupus"-type circulating anticoagulant.* **Elias, M and Eldor, A.** 1984, Arch Intern Med, Vol. 144, p. 510.
2. *Anticardiolipin antibodies: Detection by Radioimmunoassay and Association with Thrombosis in Systemic Lupus Erythematosus.* **Harris, E N, et al.** January 1, 1983, Lancet, p. 1211.
3. *Anti-phospholipid antibodies in Connective Tissue Diseases: Their Relation to the Anti-phospholipid Syndrome and forme fruste Disease.* **Buchanan, R RC, et al.** 1989, J Rheumatol, Vol. 16, p. 757.
4. *The Anticardiolipin Syndrome.* **Hughes, G RV, Harris, E N and Gharavi, A E.** 3, 1986, J Rheumatol, Vol. 13, p. 486.
5. *Anticardiolipin Antibodies- Clinical Associations.* **Asherson, R A and Harris, E N.** 1986, Postgraduate Med J, Vol. 62, p. 1081.
6. **Asherson, R A.** Antibodies to Phospholipid and the "Anti-phospholipid Syndrome." *Immunology Updates.* 1989, Vol. 1, 2, p. 1.
7. *Anti-phospholipid Antibodies.* **Harris, E N, Gharavi, A E and Hughes, G RV.** 3, 1985, Clin Rheum Dis, Vol. 11, p. 591.

8. *The Prevalence of Anti-cardiolipin Antibodies in a Healthy Elderly Population and its Association with Antinuclear Antibodies.* **Fields, R A, et al.** 1989, *J Rheumatol*, Vol. 16, p. 623.
9. *Evaluation of the anti-cardiolipin test: report of an international workshop held 4 April 1986.* **Harris, E N, et al.** 1987, *Clin Exp Immunol*, Vol. 68, p. 215.
10. **US Centers for Disease Control.** *Manual Guide – Safety Management No. CDC–22 Decontamination of Laboratory Sink Drains to Remove Azide Salts.* Atlanta : Centers for Disease Control, 1976.
11. —. *HHS Publication No. (CDC) 93-8395, 3rd ed: Biosafety in Microbiological and Biomedical Laboratories.* Washington DC : US Government Printing Office, 1993.
12. **US Department of Labor, Occupational Safety and Health Administration.** *29 CFR Part 1910.1030, Occupational safety and health standards, bloodborne pathogens.*
13. **US Department of Health and Human Services.** *HHS Publication No. (CDC) 21-11: Biosafety in Microbiological and Biomedical Laboratories. 5th ed.* Washington DC : US Government Printing Office, 2009.
14. **World Health Organization.** *Laboratory Biosafety Manual 3rd ed.* Geneva : World Health Organization, 1991.
15. **Package Insert for Calibration Samples.** From the Antiphospholipid Standardization Laboratory, University .
16. *Solid Phase Anti-cardiolipin Test Revisited.* **Harris, E N.** 1988, *Am J Med*, Vol. 85, p. 599.
17. **Moore, J E and Mohr, C F.** Biologically False Positive Serologic Tests for Syphilis: Type, Incidence, and Cause. *JAMA.* 1952, Vol. 150, 5, p. 467.

QUICK REFERENCE PROCEDURE

CARDIOLIPIN ANTIBODY (IGG) TEST

- Prepare Wash Buffer from Wash Concentrate.
- Dilute each control and specimen 1:100 in Specimen Diluent.
- Add 100 μ L of Specimen Diluent into the first well as a substrate blank.
- Pipet 100 μ L of the prediluted calibrators and the diluted controls and specimens into coated microwells and incubate 60 min at room temperature.
- Aspirate microwells and wash microwells three times with Wash Buffer.
- Pipet 100 μ L of Conjugate into microwells and incubate 30 min at room temperature.
- Aspirate microwells and wash microwells three times with Wash Buffer.
- Prepare fresh Color Developer.
- Pipet 100 μ L of Color Developer into microwells and incubate 30 min at room temperature.
- Pipet 100 μ L of Stop Solution into microwells and read results at 405 nm.

To place an order for ImmunoWELL products, contact your local distributor, or call GenBio directly for the distributor nearest you and for additional product information.

For assistance, please call toll-free 800-288-4368.

ABTS is a registered trademark of Boehringer Mannheim GmbH



15222-A Avenue of Science
San Diego, CA 92128



EMERGO EUROPE

Prinsessegracht 20
2514 AP The Hague
The Netherlands